

Product PN	FJ13ANCCA002DD01, FJ13ANCCA004FD01, FJ13ANCCA008ED01, FJ13ANCCA012CD01, FJ13ANCCA050PD01, FJ13ANCCA002DH01, FJ13ANCCA004FH01, FJ13ANCCA008EH01, FJ13ANCCA012CH01, FJ13ANCCA050PH01	Mod. 984 Rev. 02
Description	13 mm ABLUO Syringe Filters w/Cellulose Acetate Membrane	



13 mm ABLUO, Cellulose Acetate (CA)



PRODUCT DESCRIPTION

Non-sterile 13 mm syringe filter made of Acrylic housing, assembled with various pore sizes of Cellulose Acetate (CA) Membrane

Membrane	Pore Size (um)	End Fitting	Housing Material	Color	Packaging	Item Number
Cellulose Acetate (CA)	0.22	FLL/MLL	Acrylic	Blue	500/pk	FJ13ANCCA002DD01
Cellulose Acetate (CA)	0.45	FLL/MLL	Acrylic	Yellow	500/pk	FJ13ANCCA004FD01
Cellulose Acetate (CA)	0.80	FLL/MLL	Acrylic	Green	500/pk	FJ13ANCCA008ED01
Cellulose Acetate (CA)	1.20	FLL/MLL	Acrylic	Red	500/pk	FJ13ANCCA012CD01
Cellulose Acetate (CA)	5.00	FLL/MLL	Acrylic	Brown	500/pk	FJ13ANCCA050PD01
Cellulose Acetate (CA)	0.22	FLL/MLL	Acrylic	Blue	100/pk	FJ13ANCCA002DH01
Cellulose Acetate (CA)	0.45	FLL/MLL	Acrylic	Yellow	100/pk	FJ13ANCCA004FH01
Cellulose Acetate (CA)	0.80	FLL/MLL	Acrylic	Green	100/pk	FJ13ANCCA008EH01
Cellulose Acetate (CA)	1.20	FLL/MLL	Acrylic	Red	100/pk	FJ13ANCCA012CH01
Cellulose Acetate (CA)	5.00	FLL/MLL	Acrylic	Brown	100/pk	FJ13ANCCA050PH01

MANUFACTURER NAME

GVS North America
63 Community Drive
Sanford, Me 04073
 Phone: +1.866.736.1250
 eMail: CustomerCareGVSLs@gvs.com - Website: www.gvs.com

INTENDED USE / APPLICATION

Applications

- Filtration of Aqueous, Organic and Alcohol Solutions
- Analytical Sample Preparation
- IC Chromatography
- Fuel Hydraulic Fluids and Machined Parts
- Clarification
- Protein Chemistry
- Cell Culture

MATERIALS

Filter media: Cellulose Acetate

Frame/Housing Polymer: Acrylic

Color: Varies depending on pore size



Other insert(s): N/A

Regulatory Documentation Required:

PRODUCT SPECIFICATION

Product PN	FJ13ANCCA002DD01, FJ13ANCCA004FD01, FJ13ANCCA008ED01, FJ13ANCCA012CD01, FJ13ANCCA050PD01, FJ13ANCCA002DH01, FJ13ANCCA004FH01, FJ13ANCCA008EH01, FJ13ANCCA012CH01, FJ13ANCCA050PH01	Mod. 984 Rev. 02
Description	13 mm ABLUO Syringe Filters w/Cellulose Acetate Membrane	

	☒ Biocompatibility according ISO 10993-1 ☐ IMDS ☐ DEHP plasticizer Free and latex free ☒ Rohs, Directive 2002/32/CE ☐ Aging ☒ BSE/TSE, directive 2003/32/CE ☒ 1907/2006/CE (hazardous substances regulation) ☐ Dir. 67/548/CE and Reg. 1272/2008/CE (medical sector dangerous substances) ☒ Conflict minerals
--	---

PRODUCT CHARACTERISTIC	Membrane Diameter: 13 mm Effective Filtration Area: 0.76 cm ² Housing Diameter: 18 mm Housing Materials: Acrylic Inlet / Outlet: FLL / MLL Holdup Volume: <50 microliter Maximum Operating Temperature: Acrylic Abluo 50°C /122°F Maximum Operating Pressure: 80 psi Sterile: No
PRODUCT SHELF LIFE	When stored under normal storage conditions, this product should be stable for 5 years
STERILIZATION	☐ EtO ☐ Gamma ☐ Beta ☐ Steam ☐ e-beam ☒ Not Required
COMPLIANCE	The Quality management system is in compliance with ISO 9001:2008, ISO/TS 16949:2009
DRAWING	  1. Female Luer Lock Inlet (ISO 594) 2. Male Luer Lock Outlet (ISO 594) 3. CA Membrane Dimensions in mm, for ref. only

PRODUCT SPECIFICATION



Product PN	FJ13ANCCA002DD01, FJ13ANCCA004FD01, FJ13ANCCA008ED01, FJ13ANCCA012CD01, FJ13ANCCA050PD01, FJ13ANCCA002DH01, FJ13ANCCA004FH01, FJ13ANCCA008EH01, FJ13ANCCA012CH01, FJ13ANCCA050PH01	Mod. 984 Rev. 02
Description	13 mm ABLUO Syringe Filters w/Cellulose Acetate Membrane	

<

PRODUCT SPECIFICATION

Product PN	FJ13ANCCA002DD01, FJ13ANCCA004FD01, FJ13ANCCA008ED01, FJ13ANCCA012CD01, FJ13ANCCA050PD01, FJ13ANCCA002DH01, FJ13ANCCA004FH01, FJ13ANCCA008EH01, FJ13ANCCA012CH01, FJ13ANCCA050PH01	Mod. 984 Rev. 02
Description	13 mm ABLUO Syringe Filters w/Cellulose Acetate Membrane	

REVISIONS AND APPROVALS:

DATE	REV.	REASON FOR CHANGE	ISSUED AND CONTROLLED BY: (name /function and signature)	APPROVED BY: (name /function and signature)
3/31/17	0	Initial Release	Joe DeSisto, Director, Process Engineering 	Kevin Wrigley, Director, Quality 
09/25/2019	1	Updated information for 100/PK	Joe DeSisto, Director, Process Engineering 	Kevin Wrigley, Director, Quality 

CustomerApproval:

We accept this material specification as a part of the agreed terms of delivery

Company name _____

Approved by: _____
(Name, Function) (Signature)

Date _____
(Company stamp)

Please send back this document signed for approval. If we will not receive this specification signed , we consider the first order placed as implicit approval.